

M/s), 10 to 20 m/s, 15% of reading ±40 FPM (±0.2 m/s); 20 to 25 m/s; ±1.5% of reading ±60 FPM (±0.3 m/s); Temperature: ±0.54°F @ 77°F (±0.3°C @ 25°C); Relative humidity: ±2% @ 77°F (25°C) (10 to 90% RH); ±4% (0 to 10% RH and 90 to 100%). Resolution: 1 FPM (0.001 m/s).

TEMPERATURE

MODEL TM99A and SERIES 1075 Range: 0 to 200°F (-20 to 100°C). Accuracy*: ±0.5% of reading + 2.0°F (±0.5% of reading + 1.0°C). Resolution: 0.1°F (0.1°C).

- *NIST calibrated.

Note: See website for more detailed specifications. Note: Instruments meet or exceed NEBB requirements.

USA: California Proposition 65

AWARNING: Cancer and Reproductive Harm - www.P65Warnings.ca.gov

(0.1 kPa).